

ePM 10M/12M

Patient Monitor

Data Sheet



Physical Specifications

Weight	ePM 10M: 4.0 Kg ePM 12M: 4.8 Kg (Standard configuration, excluding modules, recorder, battery and accessories.)
Size	ePM 10M: 269 x252 x159mm ePM 12M: 310 x289 x169mm
Display screen	Capacitive screen, support multi-touch operation. ePM 10M: 10.1-inch, 1280 x 800 pixels ePM 12M: 12.1-inch, 1280 x 800 pixels
Display channel	ePM 10M: Up to 8 waveform channels ePM 12M: Up to 10 waveform channels

ECG

Meet standards of IEC 60601-2-27 and IEC 60601-2-25.

Lead set	3-lead: I, II, III 5-lead: I, II, III, aVR, aVL, aVF, V ** 6-lead: I, II, III, aVR, aVL, aVF, Va, Vb 12-lead: I, II, III, aVR, aVL, aVF, V1 to V6 Automatic 3/5/6/12 - lead recognition.
Input signal range	± 10 mV (p-p)
Electrode offset potential tolerance	± 800 mV
Sweep speed	6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
Gain	x 0.125, x 0.25, x 0.5, x 1, x 2, x 4, auto
Waveform format	Standard, Cabrera
Bandwidth	Diagnostic mode: 0.05 to 150 Hz Monitor mode: 0.5 to 40 Hz Surgical mode: 1 to 20 Hz ST mode: 0.05 to 40 Hz
CMRR	Diagnostic mode: > 90 dB Monitor, Surgical, ST mode: > 105 dB
Pace Detection	Amplitude: ± 2 mV to ± 700 mV Width: 0.1 to 2 ms Rise time: 10 to 100 µs
Defib. protection	Withstand 5000V (360J) defibrillation
Recovery time	<5 s
Provides glasgow resting	12-lead ECG algorithm.

Heart Rate

HR rang	Adult: 15 to 300 bpm Pediatric/Neonate: 15 to 350 bpm
HR accuracy	± 1 bpm or ± 1%, whichever is greater.
HR resolution	1 bpm

Arrhythmia Analysis

Intended use for adult, pediatric and neonate.

Multi-lead, 25 classifications. Asystole, VFib/VTac, Vtac, Vent. Brady, Extreme Tachy, Extreme Brady, Vrhythm, PVCs/min, Pauses/min, Couplet, Bigeminy, Trigeminy, R on T, Run PVCs, PVC, Tachy, Brady, Missed Beats, PNP, PNC, Multif. PVC, Nonsus. Vtac, Pause, Irr. Rhythm., Afib (for adult only).

ST Segment Analysis

Intended use for adult, pediatric and neonate.

ST range	- 2.5 to + 2.5 mV
ST accuracy	± 0.02 mV or ± 10%, whichever is greater (- 0.8 to + 0.8 mV)
ST resolution	0.01 mV

QT Analysis

Intended use for adult, pediatric, and neonate.

Parameters	QT, QTc, ΔQTc
QTc formula	Bazett, Fridericia, Framingham, or Hodges
QT/QTc range	200 to 800 ms
QT accuracy	± 30 ms
QT resolution	4 ms
QTc resolution	1 ms
QT-HR range	Adult: 15 to 150 bpm Pediatric/Neonate: 15 to 180 bpm

Respiration

Lead	I or II, auto
RR range	0 to 200 rpm
RR accuracy	± 1 rpm (0 to 120 rpm) ± 2 rpm (121 to 200 rpm)
RR resolution	1 rpm
Sweep speed	3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
Apnea time	10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s

SpO₂

Meet standards of ISO 80601-2-61.

SpO ₂ module	Mindray SpO ₂ , Nellcor SpO ₂
SpO ₂ range	0 to 100 %
SpO ₂ accuracy	Adult/Pediatric: ± 2 % (70 to 100%) Neonate: ± 3 % (70 to 100%)
Perfusion indicator (PI)	Yes, for Mindray SpO ₂
Pitch tone	Yes
Dual-SpO ₂	Yes, SpO ₂ , SpO ₂ b, ΔSpO ₂
Refreshing rate	≤ 1 s

PR

PR range	20 to 300 bpm (from SpO ₂) 20 to 350 bpm (from IBP) 30 to 300 bpm (from NIBP)
PR accuracy	± 3 bpm (20 to 300 bpm, from Mindray SpO ₂) ± 3 bpm (20 to 300 bpm, from Nellcor SpO ₂) ± 1 bpm or ± 1 %, whichever is greater (from IBP) ± 3 bpm or ± 3 %, whichever is greater (from NIBP)
Refreshing rate	≤ 1 s

Temperature

Meet standard of ISO 80601-2-56.

Technique	Thermal resistance
Channels	2 channels
Temp range	0 to 50 °C (32 to 122 °F)
Temp accuracy	± 0.1 °C or ± 0.2 °F (without probe)
Temp resolution	0.1 °C
Refreshing rate	≤ 1 s

NIBP

Meet standards of ISO 80601-2-30.

Technique	Oscillometry
Operation mode	Manual, Auto, STAT, Sequence
Parameters	Systolic, diastolic, mean
Max measurement time	Adult/Pediatric: 180 s, Neonate: 90 s
Systolic range	Adult: 25 to 290 mmHg Pediatric: 25 to 240 mmHg Neonate: 25 to 140 mmHg
Diastolic range	Adult: 10 to 250 mmHg Pediatric: 10 to 200 mmHg Neonate: 10 to 115 mmHg
Mean range	Adult: 15 to 260 mmHg Pediatric: 15 to 215 mmHg Neonate: 15 to 125 mmHg
NIBP accuracy	Max mean error: ± 5 mmHg Max standard deviation: 8 mmHg
NIBP resolution	1 mmHg
Assisting venous puncture	Yes

IBP

Meet standard of IEC 60601-2-34.

Channels	Up to 4 channels
Sensitivity	5 µV/V/mmHg
Impedance range	300 to 3000 Ω
IBP range	-50 to 360 mmHg
IBP accuracy	±1 mmHg or ± 2 %, whichever is greater
IBP resolution	1 mmHg
PPV range	0 to 50 %

PAWP	Yes.
ICP measurement	Support
Support waveforms overlapping.	
C.O.	
Technique	Thermodilution
C.O. range	0.1 to 20 L/min
C.O. accuracy	±0.1 L/min or ±5%, whichever is greater
C.O. resolution	0.1 L/min
TB range	23 to 43 °C
TI range	0 to 27 °C
TB, TI accuracy	± 0.1 °C (without sensor)
TB, TI resolution	0.1 °C
Artema Sidestream CO₂	
Meet standard of ISO 80601-2-55.	
**Options: Paramagnetic O ₂ sensor.	
CO ₂ sample flow rate	120 ml/min (DRYLINE II™ watertrap for adult/pediatric) 90/70 ml/min (DRYLINE II™ watertrap for neonate)
CO ₂ sample flow rate accuracy	± 15 ml/min or ± 15 %, whichever is greater.
CO ₂ Response time	≤ 5.0 s @ 120ml/min (for adult/pediatric) ≤ 4.5 s @ 90 ml/min (for neonate) ≤ 5.0 s @ 70 ml/min (for neonate)
O ₂ Response time	≤ 5.0 s @ 120 ml/min ≤ 4.5 s @ 90ml/min
Sweep speed	3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
CO ₂ range	0 to 150mmHg
CO ₂ accuracy	Full accuracy mode: 0 - 40 mmHg: ± 2 mmHg 41 - 76 mmHg: ± 5% of reading 77 - 150 mmHg: ± 10% of reading ISO accuracy mode: Add ± 2 mmHg to the full accuracy mode
CO ₂ resolution	1 mmHg
O ₂ range	0 to 100 %
O ₂ accuracy	± 1 % (0 to 25 %) ± 2 % (25.1 to 80 %) ± 3 % (80.1 to 100 %)
O ₂ resolution	0.1 %
awRR range	0 to 150 rpm
awRR accuracy	± 1 rpm (0 to 60 rpm) ± 2 rpm (61 to 150 rpm)
Apnea time	10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s
Oridion Microstream CO₂	
Meet standard of ISO 80601-2-55.	
Sample flow rate	50 ^{-7.5} ₊₁₅ ml/min
Initialization time	30 s (typical)
Response time	2.9 s (typical)
Sweep speed	3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
CO ₂ range	0 to 150 mmHg
CO ₂ accuracy	±2 mmHg (0 to 38 mmHg) ± 5 % of the reading (0.08 % increased in error for every 1 mmHg if the reading is more than 38mmHg) (39 to 99 mmHg)
awRR range	0 to 150 rpm
awRR accuracy	±1 rpm (0 to 70 rpm) ± 2 rpm (71 to 120 rpm) ± 3 rpm (121 to 150 rpm)
Apnea time	10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s
Capnostat Mainstream CO₂	
Meet standard of ISO 80601-2-55.	
Rise time	< 60 ms
Sweep speed	3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s

CO ₂ range	0 to 150 mmHg
CO ₂ accuracy	±2 mmHg (0 to 40 mmHg) ±5 % of the reading (41 to 70 mmHg) ±8 % of the reading (71 to 100 mmHg) ±10 % of the reading (101 to 150 mmHg)
awRR range	0 to 150 rpm
awRR accuracy	±1 rpm
Multi-gas	
Meet standard of ISO 80601-2-55.	
Technique	Infrared absorption, paramagnetic properties for O ₂ monitoring
Gas	CO ₂ , O ₂ , N ₂ O, Des, Iso, Enf, Hal, Sev
Warm-up time	ISO accuracy mode: 45 s Full accuracy mode: 10 min
Sample flow rate (with DRYLINE II™ watertrap)	Adult/pediatric watertrap: 200 ml/min Neonate watertrap: 120 ml/min
Sample flow rate accuracy	± 10 ml/min or ± 10%, whichever is greater.
Delay time	< 4 s
Response time	DRYLINE II™ watertrap for adult/pediatric, 200 ml/min: CO ₂ : ≤ 4.2 s N ₂ O: ≤ 4.3 s Enf/Iso/Hal/Sev/Des: ≤ 4.5 s O ₂ : ≤ 4 s DRYLINE II™ watertrap for neonate, 120 ml/min: CO ₂ : ≤ 4 s N ₂ O: ≤ 4.2 s O ₂ : ≤ 4 s Enf/Iso/Hal/Sev/Des: ≤ 4.4 s
CO ₂ range	0 to 30 %
CO ₂ accuracy	±0.1%ABS (0 to 1%) ±0.2%ABS (1 to 5%) ±0.3%ABS (5 to 7%) ±0.5%ABS (7 to 10%)
O ₂ range	0 to 100 %
O ₂ accuracy	±1%ABS (0 to 25%REL) ±2%ABS (25 to 80%REL) ±3%ABS (80 to 100%REL)
N ₂ O range	0 to 100 %
N ₂ O accuracy	±2%ABS (0 to 20%REL) ±3%ABS (20 to 100%REL)
Enf/Iso/Hal/Sev/Des range	0 to 30 %
awRR range	2 to 100 rpm
awRR accuracy	±1 rpm (2 to 60 rpm)
Apnea time	10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s
Provide MAC value (support calibrated by age). Support two mixed gas identify and monitoring.	
BISx/BISx4	
Meet standard of IEC 60601-2-26.	
Technique	Bispectral Index
Impedance range	>5 MΩ
EEG bandwidth	0.25 to 100 Hz
BIS range	0 to 100 (BIS, BIS L, BIS R)
SQI range	0 to 100 % (SQI, SQI L, SQI R)
ASYM	0 to 100%
DSA trend	Yes
Data Review	
For 2G storage	
Trends data	Up to 120 hours @ 1min
Events	Up to 1000 events, including parameter alarms, arrhythmia events technical alarms, and so on.
NIBP	Up to 1000 sets
Full disclosure	48 hours at Maximum. The specific storage time depends on the waveforms stored and

For 16G storage	the number of stored waveforms.
Trends data	Up to 240 hours @ 1min, 2400 hours @ 10 min
Events	Up to 2000 events, including parameter alarms, arrhythmia events technical alarms, and so on.
NIBP	Up to 3000 sets
Full disclosure	48 hours for all parameter waveforms.
For 2G & 16G storage	
Interpretation of resting	20 sets of 12-lead ECG results
OxyCRG	400 OxyCRG events
ST review	Up to 120 hours @ 1 min
Minitrend	Yes

Alarms

Audible indicator	Yes, 3 different alarm tones, and prompt tone
Visible indicator	Red/yellow/cyan LED, and alarm message display

Provide AlarmSight infographic alarm indicator.

Special Functions

Clinical Assistive Application (CAA): ST Graphic™, EWS, GCS, 24h ECG summary, NIBP analysis.

Calculations (drug, hemodynamic, Oxygenation, Ventilation, Renal), and Titration table.

Wi-Fi Communications

Protocol	IEEE 802.11a/b/g/n
Modulation mode	DSSS and OFDM
Operating frequency	IEEE 802.11b/g/n (2.4G): ETSI/FCC/KC: 2.4 to 2.483 GHz MIC: 2.4 to 2.495 GHz IEEE 802.11a/n (5G): ETSI: 5.15 to 5.35 GHz, 5.47 to 5.725 GHz FCC: 5.15 to 5.35 GHz, 5.725 to 5.82 GHz MIC: 5.15 to 5.35 GHz KC: 5.15 to 5.35 GHz, 5.47 to 5.725 GHz, 5.725 to 5.82 GHz
Channel spacing	5 MHz @ 2.4 GHz, 20 MHz @ 5 GHz
Wireless baud rate	IEEE 802.11a: 6 to 54 Mbps IEEE 802.11b: 1 to 11 Mbps IEEE 802.11g: 6 to 54 Mbps IEEE 802.11n: 6.5 to 72.2 Mbps
Output power	< 20dBm (CE requirement: detection mode- RMS) < 30dBm (FCC requirement: detection mode- peak power)
Operating mode	Infrastructure
Data security	WPA-PSK, WPA2-PSK, WPA-Enterprise, WPA2-Enterprise (EAP-FAST, EAP-TLS, EAP-TTLS, PEAP-GTC, PEAP-MSCHAPv2, PEAP-TLS, LEAP) Encryption: TKIP and AES

Interfacing

Main unit	AC power connector (1) VGA port (1) Network connector (1), RJ45 USB 2.0 connector (2) Analog output/nurse call/defib. Sync. Port (1) Integrated module rack (1), for 2 slots
Barcode scanner	Support 1D and 2D barcode
Remote control	Support
Thermal recorder	3 traces (paper 50 mm width, 20 m length)
Network printer	Support
Power	
Line voltage	100 to 240 VAC (±10 %)
Maximum current	2.0A
Frequency	50/60 Hz (±3 Hz)
Battery	Rechargeable lithium-ion battery, 2600mAh/4500mAh Rechargeable smart lithium-ion battery 5600mAh ePM 10M/12M:>2 hours run time (2600mAh) ePM 10M/12M:>4 hours run time (4500mAh) ePM 10M:>6 hours run time (5600mAh x1) ePM 12M:>4.5 hours run time (5600mAh x1) ePM 12M:>9 hours run time (5600mAh x2)
Recharge time (power off)	2.5 hours to 90%(2600mAh) 5 hours to 90% (4500mAh) 5 hours to 90% (5600mAh x1) 10 hours to 90% (5600mAh x2)

Environmental requirements

Temperature	Operating: 0 to 40 °C (without AG), 10 to 40 °C (with AG) Storage: -20 to 60 °C
Humidity	Operating: 15 to 95 % (non condensing) Storage: 10 to 95 % (non condensing)
Barometric	Operating: 427.5 to 805.5 mmHg (57.0 to 107.4 kPa) Storage: 120 to 805.5 mmHg (16.0 to 107.4 kPa)

Some of functions marked with an asterisk may not be available. Please contact your local Mindray sales representative for the most current information.

Mindray Building, Keji 12th Road South,
High-tech Industrial Park, Nanshan, Shenzhen 518057, P.R. China
Tel: +86 755 8188 8998 Fax: +86 755 26582680
E-mail: intl-market@mindray.com www.mindray.com

mindray | Healthcare with reach are registered trademarks or trademarks owned by Shenzhen Mindray Bio-medical Electronics Co., LTD
© 2018 Shenzhen Mindray Bio-medical Electronics Co., Ltd. All rights reserved. Specifications subject to changes without prior notice.
P/N: ENG-ePM 10M/12M Datasheet-210285x4P-20190228

mindray